



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/080,367	02/21/2002	Andrew Li	2002P000329US	2638
7590 04/02/2004			EXAMINER	
Siemens Corporation Intellectual Property Department 186 Wood Avenue South Iselin, NJ 08830			IMAM, ALI M	
			ART UNIT	PAPER NUMBER
			3737	

DATE MAILED: 04/02/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Advisory Action

Application No.

10/080,367

Applicant(s)

LI ET AL.

Examiner

Ali Imam

Art Unit

3737

--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

THE REPLY FILED 19 March 2004 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. Therefore, further action by the applicant is required to avoid abandonment of this application. A proper reply to a final rejection under 37 CFR 1.113 may only be either: (1) a timely filed amendment which places the application in condition for allowance; (2) a timely filed Notice of Appeal (with appeal fee); or (3) a timely filed Request for Continued Examination (RCE) in compliance with 37 CFR 1.114.

PERIOD FOR REPLY [check either a) or b)]

- a) ☒ The period for reply expires 3 months from the mailing date of the final rejection.
- b) ☐ The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection. ONLY CHECK THIS BOX WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

1. ☐ A Notice of Appeal was filed on _____. Appellant's Brief must be filed within the period set forth in 37 CFR 1.192(a), or any extension thereof (37 CFR 1.191(d)), to avoid dismissal of the appeal.
2. ☐ The proposed amendment(s) will not be entered because:
- (a) ☐ they raise new issues that would require further consideration and/or search (see NOTE below);
 - (b) ☐ they raise the issue of new matter (see Note below);
 - (c) ☐ they are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or
 - (d) ☐ they present additional claims without canceling a corresponding number of finally rejected claims.

NOTE: _____

3. ☐ Applicant's reply has overcome the following rejection(s): _____.
4. ☐ Newly proposed or amended claim(s) _____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).
5. ☒ The a) ☐ affidavit, b) ☐ exhibit, or c) ☒ request for reconsideration has been considered but does NOT place the application in condition for allowance because: See Continuation Sheet.
6. ☐ The affidavit or exhibit will NOT be considered because it is not directed SOLELY to issues which were newly raised by the Examiner in the final rejection.
7. ☒ For purposes of Appeal, the proposed amendment(s) a) ☐ will not be entered or b) ☐ will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.

The status of the claim(s) is (or will be) as follows:

Claim(s) allowed: 24.Claim(s) objected to: 2,3,5-7 and 11.Claim(s) rejected: 1,4,8-10,12-23 and 25.

Claim(s) withdrawn from consideration: _____.

8. ☐ The drawing correction filed on _____ is a) ☐ approved or b) ☐ disapproved by the Examiner.
9. ☐ Note the attached Information Disclosure Statement(s) (PTO-1449) Paper No(s). _____.
10. ☐ Other: _____



Ali Imam
Primary Examiner
Art Unit: 3737

Continuation of 5. does NOT place the application in condition for allowance because: Mucci's ultrasound system and method is directed to an automatic determination of transmit power level. Particularly, the system and method takes into account the relevant noise, signal parameters, and the range of signal values/dynamic range in order to adjust and determine the transmit power level automatically (col. 3, lines 50-56). Therefore, the lowest value of the dynamic range would inherently be taken into account since Mucci teaches in col. 8, lines 17-67 that it's system and method eliminates the need for the operator to select a gray-scale function in conjunction with other system controls such as transmit power level and receive gain. Furthermore, Mucci's system and method does this operation automatically so as to fit the received data to the dynamic range of the display as well as to provide an amount of non-linearity to achieve noise rejection and signal contrast appropriate to diagnostic Purposes..